



King's Research Portal

Document Version
Peer reviewed version

[Link to publication record in King's Research Portal](#)

Citation for published version (APA):

van Engelen, A., Mariscal Harana, J., Schneider, T., Florkow, M., Charlton, P. H., Ruijsink, B., de Blik, H., Valverde Perez, I., Charakida, M., Pushparajah, K. E., Botnar, R. M., & Alastruey-Armon, J. (2017). Validation of Non-Invasive MRI-based Assessment of Central Blood Pressure in a Population of Repaired Coarctation Patients. *Circulation*, 136(Suppl 1).

Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

Take down policy

If you believe that this document breaches copyright please contact librarypure@kcl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.

Validation of non-invasive MRI-based assessment of central blood pressure in a population of repaired coarctation patients

Arna van Engelen¹, Jorge Mariscal-Harana¹, Torben Schneider², Mateusz Florkow^{1,2}, Peter Charlton¹, Bram Ruijsink¹, Hubrecht de Bliet³, Israel Valverde¹, Marietta Charakida¹, Kuberan Pushparajah¹, Rene Botnar¹, Jordi Alastruey¹

1. Division of Imaging Sciences and Biomedical Engineering, King's College London
2. Philips Healthcare, Guildford, UK
3. HSDP Clinical Platforms, Philips Healthcare, Best, the Netherlands

Introduction

Central blood pressure (CBP) is a highly prognostic cardiovascular risk factor. It can be accurately assessed invasively with catheterization, however, this is costly and has associated risks due to its invasive nature. It is therefore necessary to develop novel non-invasive techniques for accurate CBP estimation. We present the clinical validation of a new approach using Magnetic Resonance Imaging (MRI) and one-dimensional (1-D) hemodynamic modelling to non-invasively determine CBP.

Methods

CBP measurement were obtained by invasive catheterization in 8 patients (7 male, 20±9 years) following postnatal coarctation repair, directly followed by MRI. Gold standard invasive pressure was recorded at the ascending aorta over several breathing cycles. MRI scanning included: contrast-enhanced MRA of the chest (resolution 1.3±0.3 mm in-plane, 2.3±0.5mm slice thickness), and time-resolved 2D phase-contrast MRI at the level of the ascending and diaphragmatic aorta (89±46 time points, resolution 1.8±0.4 mm, thickness 7.6±0.7 mm). Arterial geometry, the ascending flow profile, and pulse wave velocity were all semi-automatically obtained from the MRI data. Those were used as inputs to a previously optimised 1-D hemodynamic model, combined with the mean and diastolic blood pressure. All analysis was performed within an integrated software prototype.

Results

The average absolute error in modelled systolic CBP was 4.9±4.2 mmHg, the error for diastolic CBP was 4.2±2.8 mmHg and for the pulse pressure 8.0±4.9 mmHg. R² correlation values between invasive and modelled pressures were 0.89 (systolic), 0.77 (diastolic) and 0.74 (pulse pressure). A more detailed analysis is shown in the Bland-Altman figures. The time required was 5-10 minutes for MRI analysis and less than 5 minutes to run the model.

Conclusions

It is possible to measure CBP non-invasively using MRI data and 1-D modelling, as shown by close agreement of this approach with invasive measurements.

Acknowledgements

This research has been supported by an EPSRC Technology Strategy Board CR&D Grant (EP/L505304/1). JMH gratefully acknowledges the support of a PhD Fellowship awarded by the King's College London and Imperial College London EPSRC Centre for Doctoral Training in Medical Imaging. This research was supported by an EPSRC Grant (EP/K031546/1), the National Institute for Health Research (NIHR) Biomedical Research Centre award to Guy's and St Thomas' NHS Foundation Trust in partnership with King's College London, and by the NIHR Healthcare Technology Co-operative for Cardiovascular Disease at Guys and St Thomas NHS Foundation Trust.

Bland-Altman Analysis

